



**BIDS AND A WARDS COMMITTEE**

**BID BULLETIN NO. 1**

Supply, Delivery, Testing & Installation of Various Technical and Scientific Equipment for Improved Analysis and Upgrading of the Different Laboratories under ILD CY 2024  
 (Per lot basis)

November 22, 2023

This Bid Bulletin is issued to clarify, modify or amend items in the Bidding Documents. Accordingly, this shall form an integral part of the said Documents.

PARTICULAR	AMENDMENTS / CLARIFICATIONS
<p>Section VII. Technical Specifications</p> <p><b>LOT 1</b>            ITEM Nos 1.- 1 unit - Kjeldahl Digestor with Scrubber and Kjeldahl Distillation Unit with Benchtop Chiller</p> <p>Distillation capacity: ~40 mL/min at 230 V (at 100% steam addition)</p>	<p>Section VII. Technical Specifications</p> <p><b>LOT 1</b>            ITEM Nos 1.- 1 unit - Kjeldahl Digestor with Scrubber and Kjeldahl Distillation Unit with Benchtop Chiller</p> <p>Distillation capacity: <b>30-40</b> mL/min at 230 V (at 100% steam addition)</p>
<p><b>LOT 3</b>            ITEM Nos 13.- 1 unit - UV-VIS Spectrophotometer            Spectral band width: 1 nm            Wavelength accuracy: ± 0.3 nm (at Xe 260.6, 484.3, 881.9 nm; Hg 253.7, 435.8, 546.1 nm)</p>	<p><b>LOT 3</b>            ITEM Nos 13.- 1 unit - UV-VIS Spectrophotometer            Spectral band width: 1 nm or <b>better</b>            Wavelength accuracy: ± 0.3 nm (at Xe 260.6, 484.3, 881.9 nm; Hg 253.7, 435.8, 546.1 nm or <b>better</b>)</p>
<p><b>LOT 14</b>            ITEM Nos. 34 - 1 unit - Gas Chromatography tandem Mass Spectrometry(GCMS)</p>	<p><b>LOT 14</b>            ITEM Nos. 34 - 1 unit - Gas Chromatography tandem Mass Spectrometry(GCMS-<b>MS</b>)</p> <p><b>Certificate of Availability of part for at least 10 years</b></p> <p><b>Equipment to be delivered is brand new</b></p> <p><b>2 years warranty on parts and services w/ free preventive maintenance within the warranty period</b></p>
<p><b>LOT 16</b>            ITEM Nos. 43 - 1 unit - Laboratory Drying Oven            Temperature range: ambient plus 5C to 300C or better            Volume: 112 Liters minimum            External dimensions: WHD 700-720 x720- 734 x 590-610 mm</p>	<p><b>LOT 16</b>            ITEM Nos. 43 - 1 unit - Laboratory Drying Oven            Temperature range: ambient plus 5C to <b>250</b> C or better            Volume: <b>100</b> Liters minimum            External dimensions: <b>WDH 650-720 x 525- 734 x 590-860</b> mm</p>



Republic of the Philippines  
**REGIONAL FIELD OFFICE NO. 5**  
San Agustin, Pili, Camarines Sur

The Secretariat is hereby directed to inform all prospective bidders for the concerned projects and post a copy of the same in PhilGeps, designated Bid Bulletin Board and in the agency website.

For guidance and information of all concerned.

**(Signed)**  
**LORENZO L. ALVINA**  
Chief, Integrated Laboratory Division  
BAC Chairperson